ADVANCED INSTRUMENTS SAS

FICHA TECNICA

PRODUCTO	**TERMOMETRO DIGITAL INFRAROJO SIN CONTACTO**		
MODELO	**F01**		
PAIS DE ORIGEN	CHINA	PRESENTACION	CAJA POR 1 UNIDAD
VIDA UTIL	1 AÑO	MARCA	SIN MARCA
CONDICION	NUEVO	CERTIFICADOS	CERTIFICADO DE CONFORMIDAD / TECHNICAL DATA SHEET

PARAMETROS TECNICOS			
PANTALLA	PANTALLA LCD DIGITAL	PODER DE RESOLUCION	0.1°C /0.1°F
	3 TONOS DE ILUMINACION	UINIDAD DE TEMPERATURA	°C / °F
DISTACION DE MEDICION	5 CM - 15 CM	TEMPERATURA AMBIENTE DE OPERACIÓN	5°C ~ 40°C
LOCALIZACION TOMA DE TEMPERATURA	FRENTE Y SUPERFICIE DE OBJETOS	PRECISION	±0.2°C /±0.4°F
TAMAÑO	160 x 100 x 40 mm	FUENTE DE ENERGIA	2 PILAS AAA (NO INCLUIDAS)
PESO	100 g	VELOCIDAD DE MEDICION	0.5 SEGUNDOS
APAGADO AUTOMATICO	20 SEGUNDOS	ALARMA	EN TEMPERTURA > 38°C
GARANTIA	6 MESES	RANGO DE MEDICION	32°C ~ 42.9°C
ACCESORIOS	MANUAL DE INSTRUCCIONES EN INGLES	USOS	TEMPERATURA CORPORAL, TEMPERATURA DE LIQUIDOS (AGUA, LECHE)
ALMACENAMIENTO	TEMPERATURA DE ALMACENAMIENTO: ~20°C ~ 55°C		
	HUMEDAD: 10% ~ 85%		
	ESPACIO CON ADECUADA VENTILACION SIN GASES CORROSIVOS		
	EVITE GOLPES, VIBRACIONES, EXPOSICION A SALPICADURA DE AGUA O EXPOSICION DIRECTA DEL SOL DURANTE EL TRANSPORTE		

CERTIFICADOS

CELAB®
Via Maria snc
64100 Latina
Italy
celab@celab.com

celab

CERTIFICATE

Certificate Number UCN : [illegible]
Job : [illegible]
Date of issue : 2020-08-05
Certificate valid up to : 2024-08-05

Brand Name : see below
Type : Infrared Forehead Thermometers
Model N : [illegible]

Manufacturer : SHENZHEN HONGYI BIOTECHNOLOGY CO.,LTD
Address : A-202, No.5, Shenyanhu Area, Pingshan Town, Pingshan District, Shenzhen 518118 P.R. China

Standard Used : EN 1041:2008, EN ISO 15223-2:2010, EN ISO 14971:2012, EN ISO 15223-1:2016, EN 60601-1:2006+A1:2013, EN 60601-1-2:2015

Conclusion :
After inspection of the technical documentation issued by the customer, and in his request, we express our opinion that the product meets the requirements of the following directives and standards:
93/42/EEC Medical devices (MDD)

This opinion is only valid for the directive, the equipment and configuration described, in conjunction with the test data detailed above and with compliance to the applicable requirements for the product .
The following manufacturer documents were examined :

Presence of Declaration of conformity template	✔ OK
Presence of test report using standards as mentioned in the Declaration of conformity	✔ OK
Test report reference : MOS2003112961003	
Presence of CE symbol in the product. Size:	✔ OK
Presence of instruction manual	✔ OK
Use of valid Harmonized standard in the declaration of conformity	✔ OK
Presence of product description in the technical construction file	✔ OK

Copyright of this Certificate is owned by CELAB® Italy and may not be reproduced otherwise than in full and with the prior approval of the General Manager. Use of this certificate is subjected to Celab regulation available on Celab web site.
Check the authenticity of this certificate and related information before use on the web site www.celab.com introducing the UCN number in the 'Check document authenticity' area. You will see copy of this certificate and regulation on certificate use. This document is released only for scope allowed by laws- Do not use this document without full understanding of regulation.

Massimiliano Bertoldi
General Manager – CELAB
www.celab.com

Doc 121 Voluntary Certificate rev 3.32

Certificate of conformity
合格证

Product name 产品名称: Infrared thermometer 红外额温计
Product model 产品型号: F01
Executive standard 执行标准: 粤械注准20192070842
Production Lot 批次号: E20200402001
Date of manufacture 生产日期: 2020/4/2
Examination clerk 检验员
term of validity 有效期: 5 years 5年
合格
QC02
深圳鸿益生物科技有限公司
Shenzhen Hongyi Biotechnology Co., Ltd
HUAWEI Mate 9
LEICA DUAL CAMERA

Annex : Regulation for Voluntary Certification Activities

Doc 121 Voluntary Certificate rev 3.32